

# Certificate of Analysis

## Conductivity Standard Solution

**20  $\mu$ S/cm @ 25°C**

Product No:	CSKC20	Date of Measurement:	13/11/2023
Lot No:	CSKC2023L1	Date of Sample Receipt:	13/11/2023
Expiry Date*:	28/11/2024	Date of Manufacture*:	13/11/2023

<b>Specification:</b>	<b>Mean Measured Value:</b>
19.80 - 20.20 $\mu$ S/cm @ 25°C	19.91 $\mu$ S/cm @ 25°C

### Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPCOND. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed  $\pm 1\%$  at 95% confidence level,  $k=2$ .

### Metrological Traceability:


Measurement taken by comparison with standard prepared from National Institute of Standards and Technology (USA), Standard Reference Material 999 (Potassium Chloride). Electrode used for measurement: Platinised Platinum Dip Cell. Reference: ASTM D-1125 Method A.

### Accreditation:

Reagecon Diagnostics Ltd. is accredited by the American Association for Laboratory Accreditation, under scope 6739.03, for the test method, TPCOND, used to generate the above result. This accreditation deems Reagecon competent on a quality systems level and a technical level to perform the tests on the scope of accreditation. Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPCOND is technically valid and is supported by appropriate uncertainty measurements.

Date of Issue of the Certificate :  
14/11/2023

QC Technician  
HOSZOWSKA Anna



All raw materials used to prepare this product are of high purity.  
\*The detail above is based on information supplied in writing by Reagecon Manufacturing.  
Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing  
This Certificate must not be reproduced except in full. Rev-QL001