

Certificate of Analysis

Redox Standard

124 ± 5mV @ 25°C (Pt v Ag-AgCl Ref 3M KCl)

Product No:	RS124	Date of Measurement:	22/09/2023
Lot No:	RS12423J1	Date of Sample Receipt:	22/09/2023
Expiry Date*:	28/09/2025	Date of Manufacture*:	22/09/2023

Specification:

119 - 129 mV @ 25°C

Mean Measured Value:

129 mV @ 25°C


Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. The measurement cell used had a silver - silver chloride reference and 3M KCl electrolyte to give a measuring system of: Pt | Sample | 3M KCl | AgCl | Ag. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :
22/09/2023

QC Technician
HOSZOWSKA Anna

A handwritten signature in black ink, appearing to read "Anna Hoszowska".

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL002