



Certificate of Analysis

Refractive Index Standard

Refractive Index Solvent Based Standard 1.40485 nD @ 20°C

Product No: Date of Measurement: 21/03/2023 Lot No: RI014023C1 Date of Sample Receipt: 21/03/2023 28/03/2025 Date of Manufacture*: 21/03/2023 Expiry Date*:

Mean Measured Value: Specification: 1.39985 - 1.40985 nD @ 20°C 1.40861 nD @ 20°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPRIA 01. The Refractive Index value of the standard was determined using a high-performance, calibrated, temperature-controlled refractometer @20°C using sodium light at 589.3nm. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.00014 nD (0.01%) @ 20°C at 95% confidence level, i.e., coverage factor k=2.

Metrological Traceability:

This Test Result provides traceability to high purity ISO 17034 Certified Reference Materials.

Usage Instructions:

Prior to use, the contents of the dropper bottle should be brought to room temperature and mixed thoroughly by inversion to ensure a homogeneous mix.

Remove cap and discharge 4-5 droplets of liquid to waste to clear the nozzle.

Replace cap immediately after use.
Product must be stored at 5 to 30°C when not in use.

Accreditation:

Reagecon Diagnostics Ltd. is accredited by the American Association for Laboratory Accreditation, under scope 6739.03, for the test method, TPRIA 01, used to generate the above result. This accreditation deems Reagecon competent on a quality systems level and a technical level to perform the tests on the scope of accreditation. Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPRIA 01 is technically valid and is supported by appropriate uncertainty measurements.

Date of Issue of the Certificate:

21/03/2023

QC Technician

HOSZOWSKA Anna

Ann Horaval

All raw materials used to prepare this product are of high purity. *The detail above is based on information supplied in writing by Reagecon Manufacturing. Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing This Certificate must not be reproduced except in full. Rev-QL003