

CERTIFICATE

Aqueous calibration solution ASTASOL[®]

This Certificate is designed in accordance with ISO Guide 31

Category: **Certified reference material**

Analytes: **Br⁻, Cl⁻, F⁻, NO₃⁻, PO₄³⁻, SO₄²⁻**

Product code: **AN 9102 (MH) IC**

Starting primary compounds and theirs purities (%):

KBr 99.997; KCl 99.998; NaF 99.995; NH₄NO₃ 99.999; NH₄H₂PO₄ 99.999, (NH₄)₂SO₄ 99.999

Matrix:

Ultrapure demineralized water (resistivity $\geq 18 \text{ M}\Omega\cdot\text{cm}$, 0.22 μm filtered).

The solution is filtered through a membrane ultrafilter 0.45 μm to avoid clogging of a chromatography column.

Density and its expanded uncertainty (k = 2): $0.9989 \pm 0.0005 \text{ g/cm}^3$ (at 20 °C)

Certified value of concentration and its expanded uncertainty (k = 2) at 20 °C

$100.0 \pm 0.2 \text{ mg/l}$ (each analyte)

$100.1 \pm 0.3 \text{ mg/kg}^*$ (each analyte)

*Mass fraction in mg/kg is derived from density

Specification:

Batch No.: 0005

The date of production: 21.04.2020

Shelf life: 3 years from the date of production

The date of first opening of the bottle:

Expiry date:.....12 months from the first opening of the bottle within shelf life period, which should be indicated on the label of the bottle as well.

Intended use:

For calibration and validation of analytical methods analysing aqueous solutions such as atomic spectrometry (AAS, AES, ICP-OES, ICP-MS), molecular absorption spectrometry, ion chromatography and selected electroanalytical methods.

Certification and traceability:

This CRM is certified on the basis of gravimetric preparation. This procedure also ensures a direct traceability to SI unit – kg. Certified values, uncertainties and traceability were further verified by primary analytical methods (gravimetric, titrimetric) as well as by instrumental methods (AAS, AES, ICP-OES) calibrated with independent reference solutions (e.g. SRM NIST, TraceCERT, in-house solid and liquid CRMs). Analytical methods and references used are listed in the following table.

Analyte	Methods	References
Br ⁻	Argentometric titration	Analytika, CRM AN 9071(1H)
Cl ⁻	Argentometric titration	TraceCERT, Fluka 39883
F ⁻	Komplexometric titration	SRM NIST 928
NO ₃ ⁻	Gravimetric	TraceCERT, Fluka 74246
PO ₄ ³⁻	Gravimetric	TraceCERT, Fluka 38364
SO ₄ ²⁻	Gravimetric	TraceCERT, Fluka 90071

Trace impurities in bottled solution (in mg/l):

Determination of trace impurities was performed by IC chromatography. Impurity levels are supplied only for information of the user and should not be used as calibration data.

Br ⁻	Cl ⁻	F ⁻	I ⁻	NO ₂ ⁻	NO ₃ ⁻	PO ₄ ³⁻	SO ₄ ²⁻
A	A	A	<0,02	<0,02	A	A	A

< x = below detection limit A = analyte

Homogeneity and stability:

It has been demonstrated that this CRM is homogeneous and its stability is guaranteed during the whole shelf life provided the solution it kept under conditions presented below.

Storing and instruction for use:

This CRM has to be stored in the original closed bottle between 5 – 30 °C. The producer guarantees a declared shelf life and expiration time provided the CRM is properly stored and professionally handled. The temperature of the solution must be 20 ± 0.5 °C before every use. It is necessary to indicate the expiration time, which depends on the date of the first time the bottle was opened. It is important to cover the screw cap and the neck of the bottle with a parafilm layer after each opening to prevent vapour phase losses. It is not recommended to use the standard solution when the bottle contains less than 10 % of the solution. Therefore, in case of non-transparent bottle, it is important to indicate the amount of the solution used, e.g. on the label. Do not pipette from the bottle. Do not return removed aliquots to bottle.

Note:

Detailed information about the production, homogeneity, stability, coding, characterization and storing of this CRM are described in Annex to this Certificate which is its integral part.

Producer:

ANALYTIKA[®], spol. s r.o.
Department of reference materials
Ke Klíčovu 2a/816
190 00 Prague 9 – Vysočany
Czech Republic

Phone/Fax: +420 286 589 616

www.analytika.net
sales@analytika.net

Phone/Fax: +420 286 589 616

Quality management systems:

ČSN EN ISO 9001:2016
ČSN EN ISO/IEC 17025:2018
ČSN EN ISO 17034:2017



Analytika, spol. s r.o., Department of RM, Reference Materials Producer No. 7501 accredited by CAI according to ČSN EN ISO 17034:2017

Manager of Department of RM:

Head of production department:

Ing. Daniela Weissarová

Mgr. Mirka Petránková

Date of the first issue: 21.04.2020

Revision:

Date of revision:

Version: 01