

Certificate of Analysis

Chemical Oxygen Demand - COD Reagents

COD Calibration Standard 100ppm

Product No:
Lot No:
Expiry date * :

COD100
COD10021J1
28/09/2023

Mean Assay:
Date of Measurement:
Date of Sample Receipt:
Date of Manufacturing * :

100.5 ppm as COD
28/09/2021
28/09/2021
28/09/2021

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Specification:

99.0-101.0ppm

Method:

Standardised in accordance to in-house method, acidimetric.

Traceability:

This volumetric solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology, NIST SRM 84L, Potassium Hydrogen Phthalate.

Reference:

Vogels Textbook of Quantitative Inorganic Analysis. Fifth Edition.

Date of Issue of the Certificate:

28/09/2021

Eileen Malone

QC Technician **Eileen Malone**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21J32

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com