





Certificate of Analysis

pH Buffer Solution

Buffer Solution pH 10.00 ± 0.01 @ 20°C

Product No: Lot No: Expiry date: 11005 1021F1 28/06/2023

Mean pH value: Date of measurement: Specification: 9.99 - 10.01 @20°C 10.00 @20°C 09/06/2021

pH Measurement:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPPHB. Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following NIST(USA), SRM 191d-I Sodium Bicarbonate, SRM 191d-II Sodium Carbonate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate. This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed \pm 0.01pH at 95% confidence level, i.e. coverage factor k =2.

Buffer Substance:

Di-sodium Tetraborate, 1M Sodium Hydroxide.

Accreditation:

Reagecon Diagnostics Ltd. is accredited by the Irish National Accreditation Board, under scope 264T, for the test method, TPPHB, used to generate the above result. This accreditation is intended only to certify that Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPPHB is technically valid and is supported by appropriate uncertainty measurements.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of issue of the certificate:

09/06/2021

26 Shala

QC Technician Orla Sheehan

Complementary information relative to this product is available at www.reagecon.com Page 1 of 1.This Certificate must not be reproduced except in full. Rev-21E24

Reagecon Diagnostics Ltd. Shannon Free Zone, Shannon, Co. Clare, Ireland. Tel +353 61 472622, Fax: +353 61 472642 Email: sales@reagecon.ie, www.reagecon.com