



Certificate of Analysis

Turbidimetric Calibration Standard

10 NTU Ratio Turbidity Standard

Product No:	CRSR-10-500
Lot No:	CRSR1021K1
Expiry date * :	28/10/2023

Certified Value:	10.0 NTU
Date of Measurement:	20/10/2021
Date of Sample Receipt:	20/10/2021
Date of Manufacturing * :	20/10/2021

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Assay Limits:

9.5 - 10.5 NTU

Method:

Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:

The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Date of Issue of the Certificate: 20/10/2021

A handwritten signature in blue ink, reading 'Patricia de Guzman'.

QC Technician **Patricia de Guzman**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21K8

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com